

Test Report

Report Number: CSTBB21031165

Verification:



Sponsor: suzhou Virgil medical Technology Co.,Ltd

Article Name: Medical Clean Paper Wiper

Model/Size 30*40cm

Test Category Consignment Inspection

CCIC Huatongwei international inspection (Suzhou) Co., Ltd

Notices

1. Please apply for rechecking within 15 days of receiving the report if there is any objection.

2. Any erasure or without special testing seal renders the report null and void.

3. The report is only valid when signed by the persons who edited, checked and approved it.

4. The report is only responsible for the test results of the tested samples.

5. The report shall not be reproduced except in full without the written approval of the company.

Address: Room 101, Building G, Ruoshui Road 388, Suzhou, Jiangsu,

China, 512123

Tel: 0512-87657288

Fax: 0512-87657288

CCIC Huatongwei international inspection (Suzhou) Co., Ltd

Report No.CSTBB21031165


Article Name	Medical Clean Paper Wiper	Article No.	YPST21030225002
Brand	/	Model/Size	30*40cm
Sponsor	suzhou Virgil medical Technology Co.,Ltd	Test Category	Registration inspection
Manufacturer Address	NO.3 shitian Road,Yangcheng lake xiangcheng District,suzhou	Lot Batch#	20210302
Manufacturer	suzhou Virgil medical Technology Co.,Ltd	Production Date	20210329
Tested Unit	suzhou Virgil medical Technology Co.,Ltd	Number of samples	20
Received Date	2021-03-29	Test Facility	CCIC Huatongwei international inspection (Suzhou) Co., Ltd
Test Item	All items		
Test Reference	Technical requirements of 'Medical Clean Paper Wiper' provided by suzhou Virgil medical Technology Co.,Ltd		
Test Result	<p>The test items of the samples are complied with the technical requirments of 'Medical Clean Paper Wiper' provided by suzhou Virgil medical Technology Co.,Ltd.</p> <p style="text-align: right;">Date of issue: 2021- 05- 11</p>		
Remarks	<p>1) The results of the transverse tensile index test are provided by Shanghai Kuien Testing Technology Service Co., Ltd. The report number is KN-20210510-5274.</p> <p>2) The purpose of test is to conduct biological evaluation experiment on the test sample, and the results are only applicable to the test sample.</p> <p>3) “——” means not applicable, “/” means blank,</p>		

Supervisory: _____ Approved: _____ Test : _____

CCIC Huatongwei international inspection (Suzhou) Co., Ltd

Report No.CSTBB21031165

No.	Tests performed	item	Requirements	Test results	Conclusion	Memo
1	Appearance	2.1.1	The thickness should be the same, the trimming should be neat, and there should be no impurities, no black spots, etc.	Meet the requirements	Confirmed	/
		2.1.2	The appearance is neat and tidy, no lint, no flocculation, etc.	Meet the requirements	Confirmed	
		2.1.3	No abnormal smell and foreign matter.	Meet the requirements	Confirmed	
2	Size	2.2	Meet the requirements of Table 1.	Cutting size: Length: 39.30 cm Width: 29.85 cm Forming size: Length: 19.75 cm Width: 14.95 cm	Confirmed	/
3	Transverse tensile index	2.3	Should not be less than 3N·m/g.	4.70 N·m/g	Confirmed	/
4	Microbiological indicators	2.4	Products should meet the requirements of microbiological indicators in Table 2.	Total number of bacterial colonies: 10.2 CFU/g Coliform group: not detected Streptococcus hemolyticus: not detected Staphylococcus aureus: not detected	Confirmed	/
5	Biocompatibility	2.5	The product must not cause adverse irritation, allergic reactions and other damage to the skin and mucous membranes.	Meet the requirements	Confirmed	/

Article Photo

Article Describe
/
Size/Model
/

Final